

Compliance Chart

Balloon Size (mm)	Pressure (atm)																				
	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22
4.00	3.48	3.56	3.63	3.71	3.78	3.84	3.91	3.97	4.00	4.04	4.08	4.11	4.13	4.15	4.17	4.20	4.22	<i>4.25</i>	<i>4.28</i>	<i>4.31</i>	<i>4.36</i>
5.00	4.37	4.47	4.56	4.66	4.75	4.82	4.89	4.95	5.00	5.03	5.08	5.11	5.14	5.16	5.18	5.21	5.24	<i>5.28</i>	<i>5.31</i>	<i>5.35</i>	<i>5.39</i>
6.00	5.15	5.28	5.41	5.53	5.66	5.77	5.86	5.94	6.00	6.05	6.09	6.14	6.17	6.21	<i>6.25</i>	<i>6.28</i>	<i>6.33</i>	-	-	-	-
7.00	5.89	6.06	6.24	6.42	6.59	6.72	6.84	6.93	7.00	7.06	7.10	7.18	7.23	7.26	<i>7.29</i>	<i>7.35</i>	<i>7.42</i>	-	-	-	-

Nominal Pressure

Rated Burst Pressure: **Do not apply a pressure exceeding this value.** (Note) Italics; reference value