

Compliance Chart

Balloon Size (mm)	Pressure (atm)																
	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18
2.00	1.85	1.87	1.90	1.92	1.95	1.97	2.00	2.02	2.04	2.05	2.07	2.08	2.10	<i>2.11</i>	<i>2.13</i>	<i>2.14</i>	<i>2.16</i>
2.50	2.28	2.33	2.37	2.41	2.45	2.47	2.50	2.52	2.54	2.56	2.59	2.61	2.63	<i>2.66</i>	<i>2.70</i>	<i>2.73</i>	<i>2.77</i>
3.00	2.78	2.84	2.88	2.92	2.94	2.97	3.00	3.02	3.05	3.08	3.11	3.15	3.19	<i>3.23</i>	<i>3.29</i>	<i>3.36</i>	<i>3.44</i>
3.50	3.20	3.25	3.32	3.38	3.42	3.46	3.50	3.53	3.56	3.60	3.63	3.67	3.71	<i>3.75</i>	<i>3.80</i>	<i>3.86</i>	<i>3.93</i>
4.00	3.62	3.67	3.77	3.83	3.89	3.95	4.00	4.04	4.07	4.11	4.16	4.19	4.23	<i>4.27</i>	<i>4.31</i>	<i>4.37</i>	<i>4.41</i>
4.50	4.11	4.19	4.28	4.35	4.41	4.46	4.50	4.54	4.57	4.60	4.64	4.66	4.70	<i>4.74</i>	<i>4.78</i>	<i>4.83</i>	<i>4.89</i>
5.00	4.60	4.71	4.79	4.86	4.92	4.96	5.00	5.04	5.06	5.09	5.11	5.13	5.17	<i>5.20</i>	<i>5.25</i>	<i>5.30</i>	<i>5.37</i>
6.00	5.21	5.32	5.46	5.60	5.75	5.88	6.00	6.09	6.15	6.21	6.26	6.31	6.37	<i>6.42</i>	<i>6.48</i>	<i>6.54</i>	<i>6.61</i>
7.00	6.11	6.27	6.44	6.60	6.75	6.89	7.00	7.08	7.15	7.22	7.25	7.37	7.45	<i>7.54</i>	<i>7.64</i>	<i>7.76</i>	<i>7.88</i>
8.00	7.29	7.43	7.61	7.72	7.83	7.92	8.00	8.09	8.14	8.21	8.28	<i>8.35</i>	<i>8.43</i>	<i>8.52</i>	<i>8.62</i>	-	-
9.00	7.81	8.01	8.24	8.50	8.69	8.85	9.00	9.11	9.19	9.27	9.36	<i>9.45</i>	<i>9.56</i>	<i>9.66</i>	<i>9.78</i>	-	-

 Nominal Pressure

 Rated Burst Pressure: **Do not apply a pressure exceeding this value.**

(Note) Italics; reference value