

# Compliance Chart

Balloon Size (mm)	Pressure (atm)																							
	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	
4.0	3.55	3.58	3.63	3.68	3.73	3.78	3.83	3.88	3.92	3.96	4.00	4.04	4.07	4.09	4.11	4.14	4.16	4.18	4.20	4.22	4.24	4.27	4.31	
5.0	4.46	4.51	4.48	4.64	4.71	4.77	4.83	4.87	4.92	4.96	5.00	5.03	5.07	5.09	5.12	5.15	5.17	5.19	5.23	5.26	5.29	5.33	5.38	
6.0	5.29	5.37	5.46	5.54	5.63	5.70	5.77	5.84	5.89	5.95	6.00	6.04	6.07	6.10	6.14	6.17	6.21	6.24	6.27	6.32	6.36	6.45	6.53	
7.0	6.03	6.12	6.25	6.37	6.49	6.59	6.70	6.78	6.86	6.93	7.00	7.07	7.12	7.17	7.22	7.26	7.31	7.37	7.44	7.51	7.58	-	-	

Nominal Pressure

Rated Burst Pressure: **Do not apply a pressure exceeding this value.** (Note) Italics;reference value