

# Compliance Chart

Balloon Size (mm)	Pressure (atm)																						
	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24
4.00	3.55	3.58	3.63	3.68	3.73	3.78	3.83	3.88	3.92	3.96	4.00	4.04	4.07	4.09	4.11	4.14	4.16	4.18	4.2	4.22	4.24	4.27	4.31
5.00	4.43	4.44	4.51	4.58	4.65	4.71	4.78	4.84	4.89	4.94	5.00	5.04	5.07	5.11	5.14	5.17	5.19	5.22	5.25	5.27	5.31	5.34	5.39
6.00	5.25	5.32	5.42	5.51	5.6	5.68	5.77	5.83	5.89	5.95	6.00	6.05	6.08	6.12	6.15	6.18	6.21	6.25	6.28	6.32	6.36	6.42	6.48
7.00	6.03	6.12	6.25	6.37	6.49	6.59	6.7	6.78	6.86	6.93	7.00	7.07	7.12	7.17	7.22	7.26	7.31	7.37	7.44	7.51	7.58	-	-

  規定圧力 (Nominal Pressure)    
   定格破裂圧 (Rated Burst Pressure)

※ 定格破裂圧を超えて使用しないでください。 \*斜体字: 参考値を示す